



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : PT Calibrator Set

Product Code : SC084K

Lot No. : FE3981

Expiration date : 2027-11-30

Components	Quantity	Lot	Expiration Date
CAL 1	2 vials	FE3981	2027-11-30
CAL 2	2 vials	FE3981	2027-11-30
CAL 3	2 vials	FE3981	2027-11-30
CAL 4	2 vials	FE3981	2027-11-30
CAL 5	2 vials	FE3981	2027-11-30

Analytical Data	Specification	Result
CAL 1	PT% : > 90	PASS
CAL 2	PT% : 55-75	PASS
CAL 3	PT% : 40-50	PASS
CAL 4	PT% : 20-35	PASS
CAL 5	PT% : < 20	PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Approval

Quality Control Manager :

04 NOV. 2025

S. LECOURT