



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description:** BIOPHEN Plasminogen LRT  
**Product Code:** 221511  
**Lot No.:** FE29651L / FE2987  
**Expiration date:** 2027-12-24

Components	Quantity	Lot	Expiration Date
R1 : Streptokinase	3 vials	FE2965b81L / FE2987	2027-12-24
R2 : Substrate	3 vials	FE2965b91L / FE2987	2027-12-24

Analytical Data	Specification	Result
Calibration curve : reactivity, slope, linearity and accuracy on CS	A405(100%) = $0.45 \pm 0.20$ Delta A405 (100-50%) > 0.15 $R^2 \geq 0.98$ Measured values for controls in acceptance ranges	0.431 0.211 PASS PASS

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

### Quality Control Approval

Date of Approval:

21 JUL. 2025  
S. LECOURT

Quality Control Manager: