



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN™ EDOXABAN CONTROL

Product Code : 225501

Lot No. : FE31281F / FE3142

Expiration date : 2028-01-15

Components	Quantity	Lot	Expiration Date
C1	6 vials	FE3128ad1F / FE3142	2028-01-15
C2	6 vials	FE3128ae1F / FE3142	2028-01-15

Analytical Data	Specification	Result
Concentration	C1 : 100 -200 ng/ml C2 : 250 -350 ng/ml	145 ng/mL 298 ng/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Manager :

Quality Control Approve

18 AOUT 2025

S. LECOURT

