

Certificate of Analysis

5-CONTROL NAPTT Control-2
Factor IXa positive control for NAPTT
6 vials of 1 mL

REF 5D-44206B

LOT 603503

EXP 2028-01

Store at 2-8°C

Analytical Data | Specifications

Factor IXa Concentration:
(Reference Std: WHO 14/316)

14.5 mIU/mL

Acceptance Range :

[11.0 - 18.0] mIU/mL

Conclusions:

Passed:
 Refused:

Date: 2025-08-13