



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : IMIDAZOLE BUFFER
Product Code : AR021B/M/N
Lot No. : FE3165RB / FE3193
Expiration date : 2028-01-21

Components	Quantity	Lot	Expiration Date
IMIDAZOLE BUFFER	1 vial *15mL (AR021B) 3 vials *15mL (AR021M) 12 vials *15mL (AR021N)	FE3165ARRB/FE3193	2028-01-21

Analytical Data	Specification	Result
Hemoclot PC clotting assay	Delta CT (100%) with control sample \leq 5%	<2%
FVIII:C clotting assay	Delta CT (100%) with control sample \leq 5%	<1%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Approval

Quality Control Manager :

25 AOUT 2025

S. LECOURT

