



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS	
Product Description :	IMIDAZOLE BUFFER
Product Code :	AR021K/L
Lot No. :	FE3125R8
Expiration date :	2028-01-14

Components	Quantity	Lot	Expiration Date
IMIDAZOLE BUFFER	3 vials *25mL (AR021K) 12 vials *25mL (AR021L)	FE3125ARR8	2028-01-14

Analytical Data	Specification	Result
Hemoclot PC clotting assay	Delta CT (100%) with control sample \leq 5%	<2%
FVIII:C clotting assay	Delta CT (100%) with control sample \leq 5%	<1%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Approval

Quality Control Manager :

18 AOUT 2025

S. LECOURT