



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : CEPHEN 2.5
Product Code : CK512K
Lot No. : FF0228TT / FF0251
Expiration date : 2028-01-23

Components	Quantity	Lot	Expiration Date
CEPHEN 2.5	6 vials	FF0228KBTT / FF0251	2028-01-23

Analytical Data	Specification	Result
aPTT on Normal Plasma	Ratio < 1.20	1.04 on STAR 1.06 on CS 1.06 on CN

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Manager :

Quality Control Approval

25 FEV. 2026

S.MONEUS