



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** ZYMUTEST PF4

**Product Code :** RK006A

**Lot No. :** FE3032

**Expiration date :** 2028-01-01

Components	Quantity	Lot	Expiration Date
PF4 pre-coated plate	12x8 wells	FE3030	2028-01-01
PF4 Standard	3 vials	FE3042	2028-01-09
PF4 Control I	1 vial	FE3040	2028-01-09
PF4 Control II	1 vial	FE3041	2028-01-09
PF4 HRP immunoconjugate	3 vials	FE3140	2028-01-17
PF4 Sample diluent	2x50 ml	FE3031	2028-01-04
Wash solution	1x50 ml	FE0010	2028-06-15
Conjugate diluent	1x25 ml	FE0281	2028-07-10
TMB	1x25 ml	250104D01	2029-01-31
Sulfuric Acid 0,45M	1x6 ml	FE0018	2028-06-19

Analytical Data	Specification	Result
Reactivity	A450 for 8 IU/ml $\geq$ 1.70	2.04
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	< 0.100	0.108 (*)
Detection threshold	$\leq$ 0.5 IU/ml	<0.5 IU/ml

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

(\*) NCI 2746 : Accepted under waiver

Date of Approval : **Quality Control Approval**

Quality Control Manager :

18 AOUT 2025

S. LECOURT

