



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description:** ZYMUTEST tPA Antigen

**Product Code:** RK011A

**Lot No.:** FE3027

**Expiration date:** 2028-01-07

| Components                       | Quantity   | Lot       | Expiration Date |
|----------------------------------|------------|-----------|-----------------|
| Anti-(h)-tPA pre-coated plate    | 12x8 wells | FE3034    | 2028-01-07      |
| tPA standard 0                   | 1 vial     | FE3021    | 2028-12-28      |
| tPA standard 1                   | 1 vial     | FE3024    | 2028-12-28      |
| tPA standard 2                   | 1 vial     | FE3025    | 2028-12-28      |
| tPA standard 3                   | 1 vial     | FE3026    | 2028-12-28      |
| Control I                        | 1 vial     | FE3022    | 2028-12-28      |
| Control II                       | 1 vial     | FE3023    | 2028-12-28      |
| Anti-(h)-tPA HRP immunoconjugate | 3 vials    | FE3130    | 2028-01-14      |
| F-Sample diluent                 | 2x50 ml    | FD2397    | 2028-05-08      |
| Wash solution                    | 1x50 ml    | FE0011    | 2028-06-15      |
| Conjugate diluent                | 1x25 ml    | FD2352    | 2028-05-01      |
| TMB                              | 1x25 ml    | 250104D01 | 2029-01-31      |
| Sulfuric Acid 0,45M              | 1x6 ml     | FE0017    | 2028-06-19      |

| Analytical Data                | Specification                                     | Result    |
|--------------------------------|---------------------------------------------------|-----------|
| Reactivity                     | A450 for 19 ng/ml $\geq$ 1.40                     | 1.9       |
| Calibration Curve and Accuracy | Controls Measured Values within acceptance ranges | PASS      |
| Blank                          | < 0.100                                           | 0.009     |
| Detection threshold            | $\leq$ 0.5 ng/ml                                  | <0.5ng/mL |

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval:

Quality Control Manager:

Quality Control Approval

18 AOUT 2025

S. LECOURT

