HYPHEN BioMed
A Company of Sysmex Group

HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : RISTOCETIN

Product Code : AG004K

Lot No. : FE3254RK / FE3267

Expiration date : 2028-02-04

Components	Quantity	Lot	Expiration Date
RISTOCETIN	3 vials	FE3254RK / FE3267	2028-02-04

Analytical Data	Specification	Result
Reactivity (% AgregMax) at 1.2mg/ml on CS or CN series	≥ 50%	82 %
Reactivity (% AgregMax) at 1.2mg/ml on Agregometer	≥ 50%	91 %

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Manager :

Quality Control Approval

25 SEP. 2025

S. LECOURT