



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

Item Description : LIAPHEN Fibrinogen

Item Code : 120102

Lot No. : FE34221M / FE3447

Expiration date : 2028-03-03

| Components         | Quantity | Lot                 | Expiration Date |
|--------------------|----------|---------------------|-----------------|
| R1 : Latex reagent | 4 vials  | FE3422JA1M / FE3447 | 2028-03-03      |

| Analytical Data   | Specification   | Result                |
|---|---|-----------------------|
| Calibration curve : reactivity, slope, linearity and accuracy | Delta A620 (0-20 $\mu$ g/ml) $\geq$ 0.40<br>Controls measured within acceptance ranges<br>$R^2 \geq$ 0.98 | 0.57<br>PASS<br>>0.99 |
| Hook effect   | > 12 g/L (in the test dilution (1:200))   | PASS                  |

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : **Quality Control Approval**

Quality Control Manager :

09 OCT. 2025

S. LECOURT