



HYPHEN BioMed
 155 rue d'Eragny
 95000 Neuville-sur-Oise
 France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN FACTOR X

Product Code : 221705

Lot No. : FE3669

Expiration date : 2028-03-21

Components	Quantity	Lot	Expiration Date
R1 : Substrate	4 vials	FE3669	2028-03-21
R2 : RVV	4 vials	FE3669	2028-03-27
R3: Buffer	4 vials	FE3669	2028-03-24

Analytical Data	Specification	Result
Calibration curve : reactivity, linearity, slope and accuracy	$A_{405} (1:5) \geq 0.75$ $\Delta A_{405} (1:5 ; 1:10) \geq 0.35$ $R^2 \geq 0.98$ Measured values for controls in acceptance ranges	1.06 0.51 1.00 PASS
Detection Threshold	$\leq 5\%$	$< 1 \%$

Calibration Curve for information

STAR

Conc.	0	50	100	200
OD	0	0.28	0.55	1.06

	Acceptance Range	MV
NC	[74 - 94]	83
AC	[28 - 34]	32

MV = Measured value

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Approval

Quality Control Manager :

27 OCT. 2025

S. LECOURT