



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN LMWH CONTROL PLASMA  
**Product Code :** 223001  
**Lot No. :** FE3736/FE3739  
**Expiration date :** 2028-03-27

Components	Quantity	Lot	Expiration Date
C3	6 vials	FE3695BOSF/FE3728	2028-03-27
C4	6 vials	FE3696BPSG/FE3729	2028-03-27

Analytical Data	Specification	Result
Concentration	C3 : 0.70 -0.90 IU/ml C4 : 1.10 -1.30 IU/ml	0.73 IU/ml 1.18 IU/ml

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval :

Quality Control Manager :

Quality Control Approval

28 OCT. 2025

S. LECOURT