



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN UFH CONTROL PLASMA
Product Code : 223101
Lot No. : FE3636 / FE3673
Expiration date : 2028-03-18

Components	Quantity	Lot	Expiration Date
C1	6 vials	FE3634BXSA / FE3671	2028-03-18
C2	6 vials	FE3635BYSB / FE3672	2028-03-18

Analytical Data	Specification	Result
Concentration	C1 : 0.10 -0.30 IU/ml C2 : 0.40 -0.60 IU/ml	0.21 IU/ml 0.52 IU/ml

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Manager :

Quality Control Approval

16 OCT. 2025

S. LECOURT