



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN LMWH C4  
**Product Code :** 224201  
**Lot No. :** FE3696SG/FE3729  
**Expiration date :** 2028-03-27

| Components | Quantity | Lot               | Expiration Date |
|------------|----------|-------------------|-----------------|
| C4         | 12 vials | FE3696BPSG/FE3729 | 2028-03-27      |

| Analytical Data | Specification         | Result     |
|-----------------|-----------------------|------------|
| Concentration   | C4 : 1.10 -1.30 IU/ml | 1.18 IU/ml |

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval :

Quality Control Manager :

Quality Control Approval

28 OCT. 2025

S. LECOURT