



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN RIVAROXABAN CALIBRATOR LOW

Product Code : 226001

Lot No. : FE35801G / FE3601

Expiration date : 2028-03-14

Components	Quantity	Lot	Expiration Date
CAL I	4 vials	FE3580DJ1G / FE3601	2028-03-14
CAL II	4 vials	FE3580DK1G / FE3601	2028-03-14
CAL III	4 vials	FE3580DL1G / FE3601	2028-03-14

Analytical Data	Specification	Result
Concentration	CAL I : < 20 ng/ml CAL II : 40 -60 ng/ml CAL III : 80 -100 ng/ml	0 ng/mL 50 ng/mL 92 ng/mL
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Manager :

Quality Control Approval

16 OCT. 2025

S. LECOURT

