



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : Tris NaCl BSA Buffer
Product Code : AR005L
Lot No. : FE3791
Expiration date : 2028-03-31

Components	Quantity	Lot	Expiration Date
Tris NaCl BSA Buffer	4 vials (AR005L)	FE3791	2028-03-31

Analytical Data	Specification	Result
Heparin Anti IIa Assay	Delta OD (0-1 IU/mL) \geq 0.5	0.777

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **Quality Control Approval**

Quality Control Manager :

30 OCT. 2025

S. LECOURT