HYPHEN BioMed
A Sysmex Group Company

HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : HEMOCLOT QUANTI. V-L

Product Code : CK065K

Lot No. : FE37871E / FE3808

Expiration date : 2028-03-31

Components	Quantity	Lot	Expiration Date
R1: coagulation factors	3 vials	FE3787FS1E / FE3808	2028-03-31
R2 : Purified Factor Xa-PLPs	3 vials	FE3787FT1E / FE3808	2028-04-10

Analytical Data	Specification	Result
Calibration curve (manual method)	Delta CT (100-10%) ≥ 35 sec	61 sec
	CT (100%) = 32 +/- 5 sec	33 sec
	$R^2 \geq 0.98$	>0.99
	Controls measured in acceptance ranges	PASS
	Normal Plasma ≤ 10%	<1 %

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **Quality Control Approval** 04 NOV. 2025

Quality Control Manager : S. LECOURT