



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: Factor VIII:C Deficient Plasma
Product Code: DP040A/K
Lot No.: FE36382R / FE3674
Expiration date: 2028-03-17

Components	Quantity	Lot	Expiration Date
Factor VIII:C Deficient Plasma	1 vial (DP040A) 6 vials (DP040K)	FE3638JV2R / FE3674	2028-03-17

Analytical Data	Specification	Result
Calibration curve	CT (0-100%) ≥ 40 sec CT (100%) < 75 sec	83sec 57sec
Concentration FVIII	$< 1\%$	$< 1\%$
Concentration FV	$> 50\%$	68%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval:

Quality Control Manager:

Quality Control Approval

13 OCT. 2025

S. LECOURT