



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: BIOPHEN Plasminogen LRT
Product Code: 221511
Lot No.: FE38371M / FE3868
Expiration date: 2028-04-10

Components	Quantity	Lot	Expiration Date
R1 : Streptokinase	3 vials	FE3837b81M / FE3868	2028-04-10
R2 : Substrate	3 vials	FE3837b91M / FE3868	2028-04-10

Analytical Data	Specification	Result
Calibration curve : reactivity, slope, linearity and accuracy on CS	$A_{405}(100\%) = 0.45 \pm 0.20$ $\Delta A_{405} (100-50\%) > 0.15$ $R^2 \geq 0.98$ Measured values for controls in acceptance ranges	0.442 0.212 1.000 PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval:

Quality Control Approval

12 NOV. 2025

S. LECOURT

Quality Control Manager: