



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN RIVAROXABAN CONTROL PLASMA  
**Product Code :** 224501  
**Lot No. :** FE37781K / FE3779  
**Expiration date :** 2028-04-03

Components	Quantity	Lot	Expiration Date
C1	6 vials	FE3778BD1K / FE3779	2028-04-03
C2	6 vials	FE3778BE1K / FE3779	2028-04-03

Analytical Data	Specification	Result
Concentration	C1 : 50 -150 ng/ml C2 : 250 -350 ng/ml	95 ng/mL 294 ng/mL

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : **Quality Control Approval** Quality Control Manager :

04 NOV. 2025

S. LECOURT