



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : HEMOCLOT™ Protein S
Product Code : CK042K
Lot No. : FE3909SQ / FE3920
Expiration date : 2028-04-18

Components	Quantity	Lot	Expiration Date
R1: Protein S Deficient Plasma	3 vials	FE3909i3SQ / FE3920	2028-04-18
R2 : Protein S Activator	3 vials	FE3909i4SQ / FE3920	2028-04-30

Analytical Data	Specification	Result
Calibration curve on CS	Delta CT (0%-1/1) ≥ 35 sec	44.9 sec
	CT(0%) = 30-50 sec	51.1 sec*
	R ² ≥ 0.98	PASS
	Controls measured in acceptance ranges	PASS
	Normal Plasma > 60%	PASS
	Plasma Deficient ≤ 10%	PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

*NCI2833 : Accepted user waiver

Date of Approval :

Quality Control Manager :

Quality Control Approval

24 NOV. 2025

S. LECOURT

