



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN NORMAL PLASMA 2
Product Code : 223602
Lot No. : FF1431
Expiration date : 2028-05-22

Components	Quantity	Lot	Expiration Date
Normal Plasma 2	10 vials	FF1431	2028-05-22

Analytical Data	Specification	Result
INR	< 1,3	1.16
aPTT ratio	≤ 1.20	1.06
Fibrinogen	≥ 2 g/L	3.21 g/L
FVIII:C	> 50%	96%
LA DVV test / DVV confirm ratio	≤ 1.20	0.98

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **Quality Control Approval**

16 JUIN 2026

Quality Control Manager :

S.MONEUS