



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN APIXABAN CONTROL LOW
Product Code : 225201
Lot No. : FE40361L / FE4052
Expiration date : 2028-05-09

Components	Quantity	Lot	Expiration Date
CI	6 vials	FE4036DW1L / FE4052	2028-05-09
CII	6 vials	FE4036DX1L / FE4052	2028-05-09

Analytical Data	Specification	Result
Concentration	CI : 15 -35 ng/ml CII : 45 -75 ng/ml	23 ng/mL 58 ng/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2025-12-09

Quality Control Manager :

Sophie LECOURT
Technical Director

