



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN APIXABAN CONTROL
Product Code : 225301
Lot No. : FE40361K / FE4054
Expiration date : 2028-05-09

Components	Quantity	Lot	Expiration Date
C1	6 vials	FE4036DR1K / FE4054	2028-07-04
C2	6 vials	FE4036DS1K / FE4054	2028-05-09

Analytical Data	Specification	Result
Concentration	C1 : 140 -260 ng/ml C2 : 340 -460 ng/ml	178 ng/ml 402 ng/ml

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: _____ Quality Control Approval: _____ Quality Control Manager : _____

U 4 FEV. 2026

S. LECOURT