



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS	
Product Description :	IMIDAZOLE BUFFER
Product Code :	AR021K/L
Lot No. :	FF0483
Expiration date :	2028-05-19

Components	Quantity	Lot	Expiration Date
IMIDAZOLE BUFFER	3 vials *25mL (AR021K) 12 vials *25mL (AR021L)	FF0483	2028-05-19

Analytical Data	Specification	Result
Hemoclot PC clotting assay	Delta CT (100%) with control sample ≤ 5%	<1%
FVIII:C clotting assay	Delta CT (100%) with control sample ≤ 5%	<1%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **10 FEB. 2026**

Quality Control Approval

Quality Control Manager : 

S.MONEUS