



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : Factor IX Deficient Plasma
Product Code : DP050A/K
Lot No. : FE4103T7 / FE4111
Expiration date : 2028-05-16

Components	Quantity	Lot	Expiration Date
Factor IX Deficient Plasma	1 vial (DP050A) 6 vials (DP050K)	FE4103JWT7 / FE4111	2028-05-16

Analytical Data	Specification	Result
Calibration curve	CT (0%) ≥ 40 SEC CT (100%) <80 SEC	126sec 53sec
Concentration FIX	< 1%	<1%
Concentration FV	> 50 %	85%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Manager :

Quality Control Approval

09 DEC. 2025

S. LECOURT

