



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS	
Product Description : ZYMUTEST Fibrinogen	
Product Code :	RK024A
Lot No. :	FE4024
Expiration date :	2028-05-08

Components	Quantity	Lot	Expiration Date
pre-coated plate	12x8 wells	FE4141	2028-05-19
Fibrinogen Standard	3 vials	FE4022	2028-05-08
Fibrinogen Control I	1 vial	FE4021	2028-05-08
Fibrinogen Control II	1 vial	FE4023	2028-05-08
Fibrinogen HRP immunoconjugate	3 vials	FE4248	2028-06-05
B2F Sample diluent	2x50 ml	FE4096	2028-05-12
Wash solution	1x50 ml	FE3948	2029-04-16
B2F Conjugate diluent	1x25 ml	FE4097	2028-05-12
TMB	1x25 ml	250301D02	2029-03-31
Sulfuric Acid 0,45M	1x6 ml	FE3546	2029-02-26

Analytical Data	Specification	Result
Reactivity	A450 for 40 ng/ml \geq 1.45	2.04
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	< 0.150	0.025
Detection threshold	\leq 1 ng/ml	< 1 ng/ml

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Quality Control Approval

Date of Approval : **13 JAN. 2026**

Quality Control Manager : **13 JAN. 2026**

P.O.
I. CORNUEJOLS