



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN a2 ANTIPLASMIN (LRT)
Product Code : 220502
Lot No. : FF13751L / FF1373
Expiration date : 2028-06-04

Components	Quantity	Lot	Expiration Date
R1 :Plasmin	3 vials	FF1375171L / FF1373	2028-06-04
R2 : SPm-41 Substrate	3 vials	FF1375181L / FF1373	2028-06-04

Analytical Data	Specification	Result
Calibration curve : reactivity, slope, linearity and accuracy on CS	Delta A405(100-50%)> 0.05 A405/min(100%) = 0.25 ± 0.10 $R^2 \geq 0.98$ Measured values for controls in acceptance ranges	0.088 0.227 PASS PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Approval

Quality Control Manager :

09 JUL. 2026

S.MONEUS