



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN AT (Anti IIa)
Product Code : 221122
Lot No. : FF00801D / FE4276
Expiration date : 2028-06-02

Components	Quantity	Lot	Expiration Date
R1 : Bovin Thrombin	2 vials	FF0080IT1D / FE4276	2028-07-10
R2 : Substrate	2 vials	FF0080IU1D / FE4276	2028-06-06
R3: Tris heparin Buffer	2 vials	FF0080IV1D / FE4276	2028-06-02

Analytical Data	Specification	Result
Calibration curve : reactivity, linearity, slope and accuracy	Delta A405 (0-100%) ≥ 0.80 $R^2 \geq 0.98$ Measured values for controls in acceptance ranges	1.48 PASS PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : Quality Control Approval

10 MARS 2026

Quality Control Manager :

S.MONEUS