



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN FVII

Product Code : 221304

Lot No. : FF0123

Expiration date : 2028-06-23

Components	Quantity	Lot	Expiration Date
R1 : Human FX	2 vials	FF0123	2028-06-23
R2 : Calcium Thromboplasstin	2 vials	FF0123	2028-07-09
R3 : Substrate	2 vials	FF0123	2028-06-27
R4 : Buffer	4 vials	FF0123	2028-06-23

Analytical Data	Specification	Result
Calibration curve : reactivity, slope, linearity on Water-Bath	A405(200%) ≥ 0.80 A405(0%) ≤ 0.10 Delta A405(100-200%) ≥ 0.30 $R^2 \geq 0.98$	1.16 0.05 0.36 1.00
Detection Threshold	$\leq 5\%$	2%

Calibration Curve for information

Conc.	200%	100%	50%	25%
OD WB	1.156	0.793	0.479	0.285
Conc.	200%	100%	50%	25%
OD STAR	1.219	0.855	0.538	0.313

	Acceptance Range	MV
NC	[97 - 117]	109
AC	[28 - 38]	34

MV = Measured value

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Manager :

Quality Control Approval

Sophie LECOURT
Technical Director

05 FEB. 2026

S.MONEUS