



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : ZYMUTEST vWF

Product Code : RK030A

Lot No. : FE3618 / FE3971

Expiration date : 2028-06-15

Components	Quantity	Lot	Expiration Date
Anti (h) vWF pre-coated plate	12x8 wells	FE4090	2029-05-07
vWF standard	3 vials	FE3959	2029-04-19
vWF Control I	1 vial	FE4210	2029-05-23
vWF Control II	1 vial	FE4209	2029-05-23
Anti-(h)-vWF HRP immunoconjugate	3 vials	FE4088	2029-05-10
Sample diluent	2x50 ml	FE0193	2028-07-03
Wash solution	1x50 ml	FE0011	2028-06-15
Conjugate diluent	1x25 ml	FE3391	2029-02-12
TMB	1x25 ml	241004D01	2028-10-31
Sulfuric Acid 0,45M	1x6 ml	FE0017	2028-06-19

Analytical Data	Specification	Result
Reactivity	A450 for 100 % \geq 1.40	1.98
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	< 0.100	0.022
Detection threshold	\leq 5%	< 1%

Calibration Curve for information

Conc.	C	C/2	C/4	C/10	C/20
OD	2.324	1.720	1.143	0.572	0.359

	Acceptance Range	MV
CI	[65 - 87]	79
CII	[23 - 35]	30

MV = Measured value

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Manager :

Quality Control Approval

20 JAN. 2026

S. LECOURT