



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : ZYMUTEST vWF: CBA

Product Code : RK038A

Lot No. : FE3617 / FE3677

Expiration date : 2028-06-19

Components	Quantity	Lot	Expiration Date
Collagen pre-coated plate	12x8 wells	FE4008	2029-04-30
vWF standard	3 vials	FE4211	2029-05-23
vWF Control I	1 vial	FE4210	2029-05-23
vWF Control II	1 vial	FE4209	2029-05-23
Anti-(h)- vWF HRP immunoconjugate	3 vials	FE4087	2029-05-10
Sample diluent	2x50 ml	FE3655	2029-03-12
Wash solution	1x50 ml	FE3949	2029-04-16
Conjugate diluent	1x25 ml	FE3391	2029-02-12
TMB	1x25 ml	250301D02	2029-03-31
Sulfuric Acid 0,45M	1x6 ml	FE0017	2028-06-19

Analytical Data	Specification	Result
Reactivity	A450 for 100 % ≥ 1.00	1.89
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	< 0.100	0.003
Detection threshold	$\leq 5\%$	< 1

Calibration Curve for information

Conc.	C	C/2	C/4	C/10	C/20
OD	2.260	1.646	1.039	0.484	0.262

	Acceptance Range	MV
CI	[63 - 85]	74
CII	[26 - 38]	30

MV = Measured value

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Manager :

Quality Control Approval

20 JAN. 2026

S. LECOURT

F820-11/ZY/038A v10