



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN NORMAL CONTROL PLASMA

**Product Code :** 223201

**Lot No. :** FF0090TP / FF0189

**Expiration date :** 2028-07-03

Components	Quantity	Lot	Expiration Date
Normal Control	12 vials	FF0090DZTP / FF0189	2028-07-03

Analytical Data	Specification	Result
INR	0.95-1.20	PASS
aPTT LR Ratio	0.85-1.15	PASS
Fibrinogen (Activity)	2.4-3.5 g/L	PASS
Fibrinogen (Antigen)	2.4-3.5 g/L	PASS
Thrombin time Ratio	< 1.20	PASS
LA ratio	≤ 1.2	PASS
Alpha 2 Anti plasmin	75-140%	PASS
Antithrombin (Activity)	75-125%	PASS
Antithrombin (Antigen)	75-125%	PASS
Factor II	75-125%	PASS
Factor V	75-140%	PASS
Factor VII	75-125%	PASS
Factor X	75-140%	PASS
Factor VIII	75-125%	PASS
Factor IX	75-125%	PASS
Factor XI	75-140%	PASS
Factor XII	75-140%	PASS
Factor XIII	75-140%	PASS
Free Protein S (Antigen)	75-140%	PASS
Plasminogen	75-140%	PASS
Protein C Chromogenic	75-140%	PASS
Protein C (Antigen)	75-140%	PASS
Protein C Clotting	75-140%	PASS
Protein S (Activity)	75-140%	PASS
von Willebrand Factor (Antigen)	75-140%	PASS
Act. PC-r (quantitative)	≤ 10 % FVL	PASS

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval :

**Quality Control Approval**

Quality Control Manager :

27 AVR. 2026

S.MONEUS