



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN DABIGATRAN CONTROL PLASMA LOW
Product Code : 225001
Lot No. : FF03221C / FF0340
Expiration date : 2028-07-17

Components	Quantity	Lot	Expiration Date
C1	6 vials	FF0322DD1C / FF0340	2028-07-17
C2	6 vials	FF0322DE1C / FF0340	2028-07-17

Analytical Data	Specification	Result
Concentration	C1 : 15 -35 ng/ml C2 : 40 -80 ng/ml	27 ng/ml 72 ng/ml

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : Quality Control Approval Quality Control Manager :

03 MARS 2026

S.MONEUS