



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN Factor XIII
Product Code : 227005
Lot No. : FF01731N / FF0188
Expiration date : 2028-07-03

Components	Quantity	Lot	Expiration Date
R1 : Thrombin	3 vials	FF0173bs1N / FF0188	2028-07-03
R2 : Detection Reagent	3 vials	FF0173bt1N / FF0188	2028-07-03

Analytical Data	Specification	Result
Calibration curve : reactivity, slope, linearity and accuracy on CS	OD/min(100%) > 0.016 Delta OD/min (100-50%) ≥ 0.006 R ² ≥ 0.98 Measured values for controls in acceptance ranges	0.022 0.01 PASS PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **Quality Control Approval**
04 FEV. 2026
S. LECOURT

Quality Control Manager : 