



HYPHEN BioMed
 155 rue d'Eragny
 95000 Neuville-sur-Oise
 France

CERTIFICATE OF ANALYSIS

Product Description : ZYMUTEST FACTOR X

Product Code : RK033A

Lot No. : FE3221

Expiration date : 2028-07-10

Components	Quantity	Lot	Expiration Date
Anti (h) Factor X pre-coated plate	12x8 wells	FE3946	2029-04-16
Factor X standard	3 vials	FE3858	2029-04-05
Factor X Control I	1 vial	FE3860	2029-04-05
Factor X Control II	1 vial	FE3859	2029-04-05
Anti-(h) Factor X HRP immunoconjugate	3 vials	FE3990	2029-04-25
Sample diluent	2x50 ml	FE3656	2029-03-12
Wash solution	1x50 ml	FE3948	2029-04-16
Conjugate diluent	1x25 ml	FE0281	2028-07-10
TMB	1x25 ml	250104D01	2029-01-31
Sulfuric Acid 0,45M	1x6 ml	FE3546	2029-02-26

Analytical Data	Specification	Result
Reactivity	A450 for 100 % ≥ 1.40	1.60
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	< 0.100	PASS
Detection threshold	$\leq 5\%$	$< 1\%$

Calibration Curve for information

Conc.	C	C/2	C/4	C/10	C/20
OD	1.777	1.279	0.842	0.422	0.246

	Acceptance Range	MV
CI	[69 - 93] %	78%
CII	[27 - 41] %	33%

MV = Measured value

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Manager :

Quality Control Approval

11 DEC. 2025

S. LECOURT