



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN RIVAROXABAN CALIBRATOR

Product Code : 222701

Lot No. : FF05731G / FF0594

Expiration date : 2028-08-22

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FF0573BA1G / FF0594	2028-08-22
CAL 2	4 vials	FF0573BB1G / FF0594	2028-08-22
CAL 3	4 vials	FF0573BC1G / FF0594	2028-08-22

Analytical Data	Specification	Result
Concentration	CAL 1 : ≤ 50 ng/ml CAL 2 : 200 -300 ng/ml CAL 3 : 400 -500 ng/ml	0 ng/ml 250 ng/ml 449 ng/ml
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	1.00 PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Manager :

Quality Control Approval

30 MARS 2026

S.MONEUS