


 HYPHEN BioMed
 155 rue d'Eragny
 95000 Neuville-sur-Oise
 France

CERTIFICATE OF ANALYSIS

Product Description : Factor XI Deficient Plasma
Product Code : DP070A/K
Lot No. : FF05291L / FF1029
Expiration date : 2028-08-15

Components	Quantity	Lot	Expiration Date
Factor XI Deficient Plasma	1 vial (DP070A) 6 vials (DP070K)	FF0529JX1L / FF1029	2028-08-15

Analytical Data	Specification	Result
Calibration curve	CT (0%) >150 sec CT (100%) <100 sec	160sec 62sec
Concentration FXI	< 1%	<1%
Concentration FV	> 50 %	67%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Approval

Quality Control Manager :

22 AVR. 2026

S.MONEUS

