



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

### CERTIFICATE OF ANALYSIS

Item Description : LIAPHEN Fibrinogen

Item Code : 120102

Lot No. : FF05841N / FF0593

Expiration date : 2028-09-01

Components	Quantity	Lot	Expiration Date
R1 : Latex reagent	4 vials	FF0584JA1N / FF0593	2028-09-01

Analytical Data	Specification	Result
Calibration curve : reactivity, slope, linearity and accuracy	Delta A620 (0-20µg/ml) $\geq$ 0.40 Controls measured within acceptance ranges $R^2 \geq$ 0.98	0.45 PASS >0.99
Hook effect	> 12 g/L (in the test dilution (1:200))	PASS

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval :

Quality Control Approval

Quality Control Manager :

08 AVR. 2026

S.MONEUS