



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS	
Product Description :	BIOPHEN UFH CONTROL PLASMA
Product Code :	223101
Lot No. :	FF0922 / FF1103
Expiration date :	2028-09-26

Components	Quantity	Lot	Expiration Date
C1	6 vials	FF0921BXV9 / FF0927	2028-09-26
C2	6 vials	FF0913BYV6 / FF1104	2028-09-26

Analytical Data	Specification	Result
Concentration	C1 : 0.10 -0.30 IU/ml C2 : 0.40 -0.60 IU/ml	0.21 IU/ml 0.53 IU/ml

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Approval

Quality Control Manager :

05 MAI 2026

S.MONEUS