



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN RIVAROXABAN CONTROL PLASMA
Product Code : 224501
Lot No. : FF07101L / FF0733
Expiration date : 2028-09-05

Components	Quantity	Lot	Expiration Date
C1	6 vials	FF0710BD1L / FF0733	2028-09-05
C2	6 vials	FF0710BE1L / FF0733	2028-09-05

Analytical Data	Specification	Result
Concentration	C1 : 50 -150 ng/ml C2 : 250 -350 ng/ml	95 ng/mL 306 ng/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Approval

Quality Control Manager :

09 AVR. 2026

S.MONEUS