HYPHEN BioMed

A Company of Sysmex Group

HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN Calibrator Factor VIIa

Product Code : 226301

Lot No. : FF0951

Expiration date : 2028-09-29

Components	Quantity	Lot	Expiration Date
Calibrator VIIa	6 vials	FF0951	2028-09-29

Analytical Data	Specification	Result
Concentration	350-500 mIU/ml	435 mUI/mL
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 12 MAI 2026

Quality Control Approval : S.MONEUS

Quality Control Manager :