



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : IMIDAZOLE BUFFER
Product Code : AR021B/M/N
Lot No. : FF1032VD / FF1040
Expiration date : 2028-10-13

Components	Quantity	Lot	Expiration Date
IMIDAZOLE BUFFER	1 vial *15mL (AR021B) 3 vials *15mL (AR021M) 12 vials *15mL (AR021N)	FF1032ARVD / FF1040	2028-10-13

Analytical Data	Specification	Result
Hemoclot PC clotting assay	Delta CT (100%) with control sample $\leq$ 5%	<1%
FVIII:C clotting assay	Delta CT (100%) with control sample $\leq$ 5%	<1%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Approval

19 MAI 2026

S.MONEUS

Quality Control Manager :

