



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : ARGATROBAN CONTROL PLASMA
Product Code : SC035K
Lot No. : FF10691D / FF1116
Expiration date : 2028-10-20

Components	Quantity	Lot	Expiration Date
C1	6 vials	FF1069CE1D / FF1116	2028-10-20
C2	6 vials	FF1069CF1D / FF1116	2028-10-20

Analytical Data	Specification	Result
C1	0.55 - 0.80 µg/mL	0.74 µg/mL
C2	1.10 - 1.40 µg/mL	1.38 µg/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Approval

Quality Control Manager :

05 JUIN 2026

S.MONEUS

