



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN AT (Anti IIa)

Product Code : 221122

Lot No. : FF12491E / FF1294

Expiration date : 2028-11-17

Components	Quantity	Lot	Expiration Date
R1 : Bovin Thrombin	2 vials	FF1249IT1E / FF1294	2028-12-04
R2 : Substrate	2 vials	FF1249IU1E / FF1294	2028-11-17
R3: Tris heparin Buffer	2 vials	FF1249IV1E / FF1294	2028-11-24

Analytical Data	Specification	Result
Calibration curve : reactivity, linearity, slope and accuracy	Delta A405 (0-100%) ≥ 0.80 $R^2 \geq 0.98$ Measured values for controls in acceptance ranges	1.497 $R^2 \geq 0.999$ PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Approval

Quality Control Manager :

30 JUL. 2026

S.MONEUS