



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : RISTOCETIN
Product Code : AG004K
Lot No. : FF1185VT / FF1216
Expiration date : 2028-11-11

Components	Quantity	Lot	Expiration Date
RISTOCETIN	3 vials	FF1185VT / FF1216	2028-11-11

Analytical Data	Specification	Result
Reactivity (% AgregMax) at 1.2mg/ml on CS or CN series	$\geq 50\%$	78 %
Reactivity (% AgregMax) at 1.2mg/ml on Agregometer	$\geq 50\%$	90 %

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : Quality Control Approval

30 JUIN 2026

S.MONEUS

Quality Control Manager :