



HYPHEN BioMed
 155 rue d'Eragny
 95000 Neuville-sur-Oise
 France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN APIXABAN CALIBRATOR

Product Code : 226201

Lot No. : FF13651H / FF1418

Expiration date : 2028-12-02

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FF1365DO1H / FF1418	2028-12-02
CAL 2	4 vials	FF1365DP1H / FF1418	2028-12-02
CAL 3	4 vials	FF1365DQ1H / FF1418	2028-12-02


Analytical Data	Specification	Result
Concentration	CAL 1 : ≤ 50 ng/ml CAL 2 : 200 -400 ng/ml CAL 3 : 400 -600 ng/ml	0 ng/mL 314 ng/mL 525 ng/mL
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Approval

Quality Control Manager :



15 JUL. 2026

S.MONEUS