



HYPHEN BioMed
 155 rue d'Eragny
 95000 Neuville-sur-Oise
 France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN EDOXABAN CALIBRATOR
Product Code : 226501
Lot No. : FF15291H / FF1542
Expiration date : 2028-12-19

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FF1529aa1H / FF1542	2028-12-19
CAL 2	4 vials	FF1529ab1H / FF1542	2028-12-19
CAL 3	4 vials	FF1529ac1H / FF1542	2028-12-19

Analytical Data	Specification	Result
Concentration	CAL 1 : ≤ 40 ng/ml CAL 2 : 200 -300 ng/ml CAL 3 : 400 -600 ng/ml	0 ng/mL 243 ng/mL 513 ng/mL
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **Quality Control Approval**

Quality Control Manager :

27 JUL. 2026

S.MONEUS

