



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Item Description : LIAPHEN Fibrinogen

Item Code : 120102

Lot No. : FF16851P / FF1702

Expiration date : 2029-01-19

Components	Quantity	Lot	Expiration Date
R1 : Latex reagent	4 vials	FF1685JA1P / FF1702	2029-01-19

Analytical Data	Specification	Result
Calibration curve : reactivity, slope, linearity and accuracy	Delta A620 (0-20µg/ml) $\geq$ 0.40 Controls measured within acceptance ranges $R^2 \geq 0.98$	0.52 PASS >0.99
Hook effect	> 12 g/L (in the test dilution (1:200))	PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **Quality Control Approval**

12 AOUT 2026

S.MONEUS

Quality Control Manager :